

常州市回春医疗器材有限公司

CHANGZHOU HUICHUN MEDICAL EQUIPMENT CO.,LTD.

Test Report of Hypodermica Needle

According to: ISO7864:1993

LOT NO.	20131028	Report NO.	20131028-3
MFG. DATE	2013/10	Spec.	18G×1 1/2"
EXP. DATE	2018/10	Test Dates	2013/10/29
Quantity	250000pcs	Report Dates	2013/11/15

Inspection items	Guideline	Result
Chemical requirements	pH≤1.0,Deference of pH should not be more than 1.0	0.10
Biological requirements	There should be sterile	Measure up
	There should be no pyrogen	Measure up
Physical requirements	Appearance of Needle Stand: No feathered edge, burr, plastic flow, and defect;	Measure up
	Needle should be no burr	Measure up
	Needle Point should be sharp	Measure up
	Needle puncture should be conform with prescription of GB18457-2001	0.72N
	Needle doesn't separate with needle stand under 34N inverse force.	Measure up
	Needle size should be conform with prescription of GB18457-2001	Size: 1.2mm Length: 38mm
	Needle Stand Color should be conform with prescription of YY/T0296	Measure up

Conclusion: Products measure up according to ISO7864:1993.

