

常州市回春医疗器材有限公司

CHANGZHOU HUICHUN MEDICAL EQUIPMENT CO.,LTD.

Test Report of Hypodermica Needle

According to: ISO7864:1993

Test NO.	20130528	Report NO.	20130528-7
Production LOT	2013/05	Spec.	23G×1 1/4"
Sterilizing LOT	2018/05	Test Dates	2013/06/05
Quantity	100000pcs	Report Dates	2013/06/20

Inspection items	Guideline	Result
Chemical requirements	pH<1.0,Deference of pH should not be more than 1.0	0.10
Biological requirements	There should be sterile	Measure up
	There should be no pyrogen	Measure up
Physical requirements	Appearance of Needle Stand:	Measure up
	No feathered edge, burr, plastic flow, and defect;	
	Needle should be no burr	Measure up
	Needle Point should be sharp	Measure up
	Needle puncture should be conform with prescription of GB18457-2001	0.42N
	Needle doesn't separate with needle stand under 34N inverse force.	Measure up
	Needle size should be conform with prescription of GB18457-2001	Size: 0.6mm Length: 32mm
	Needle Stand Color should be conform with prescription of YY/T0296	Measure up

Conclusion: Products measure up according to ISO7864:1993.

