

## Certificate of Quality

REF

455092

LOT

A161045M

Product description:

VACUETTE® TUBE 9 ml Z Serum Clot Activator

16x100 red cap-black ring, non-ridged

Following inspections were carried out and passed:

visual

X

passed

physical

X

passed

chemical

X

passed

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

98/79/EC "Directive 98/79/EC of the European Parliament and of the Council of 27th Oct 1998 on in vitro diagnostic medical devices"

ISO 6710 "Single-use containers for venous blood specimen collection"

ISO 11137 "Sterilization of health care products - Radiation"

EN 14820 "Single-use containers for human venous blood specimen collection"

ISO 9001 "Quality management systems - Requirements"

ISO 13485 "Medical devices - Quality management systems - Requirements for regulatory purposes"

2

2018-04-07

STERILE

SAL 10<sup>-6</sup>

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2016-11-14

www.qbo.com

Austria: Greiner Bio-One GmbH, office@at.gbo.com | Belgium: Greiner Bio-One BVBA/SPRL, into@be.gbo.com | Brazil: Greiner Bio-One SAS, into@fr.gbo.com | Hungary: Greiner Bio-One Bio-One Bio-One Brazil: Greiner Bio-One Brazil: Greiner Bio-One SAS, into@fr.gbo.com | Hungary: Greiner Bio-One Hungary: Greiner Bio-One Bio-One Bio-One Brazil: Greiner Bio-One Brazil: Grei