



COAGDX01-1S v001

## Certificate of Analysis

### Product Information:

|                   |                                       |
|-------------------|---------------------------------------|
| Product name:     | Petscreen® Canine Giardia Ag Test Kit |
| Product code:     | GDX01-1S                              |
| Lot number:       | GDX011S1901                           |
| Manufacture date: | 01/2020                               |
| Expiry date:      | 01/2022                               |

### Quality Control Summary:

| Component                         | Parameter                  | Result | Comments |
|-----------------------------------|----------------------------|--------|----------|
| Assay Diluent                     | Appearance                 | PASS   | -        |
|                                   | Sterility                  | PASS   | -        |
|                                   |                            |        | -        |
| Test Device – Physical properties | Membrane flow              | PASS   | -        |
|                                   | Appearance                 | PASS   | -        |
|                                   |                            |        | -        |
| Test Device – Functionality       | Low positive               | PASS   | -        |
|                                   | Mid positive               | PASS   | -        |
|                                   | High positive              | PASS   | -        |
|                                   | Negative 1                 | PASS   | -        |
|                                   | Negative 2                 | PASS   | -        |
|                                   | Negative 3                 | PASS   | -        |
|                                   |                            |        | -        |
| Kit Components                    | Leak test - pouched device | PASS   | -        |
|                                   | Leak test - diluent bottle | PASS   | -        |

| Kit Summary                 | Result |
|-----------------------------|--------|
| Total test devices assessed | 15     |
| Total test devices rejected | 0      |
| Kit status                  | PASS   |



GLOBAL DX LIMITED  
Elmbank Business Centre, Menstrie FK11 7BU  
United Kingdom  
Tel: +44 (0)1259 230830  
www.globaldx.com  
VAT No: GB 288 1975 44

Certified by:

Sharon Shepherd

Date:

24.02.20For *in vitro* diagnostic use only