

Certificate of Analysis

Product Information:

Product name:	Petscreen® Canine Heartworm (CHW) Ag Test Kit
Product code:	GDX11-1
Lot number:	GDX1112003
Manufacture date:	OCT 2020
Expiry date:	SEP 2022

Quality Control Summary:

Component	Parameter	Result	Comments
Assay Diluent	Appearance	PASS	-
	Sterility	PASS	-
Test Device – Physical properties	Membrane flow	PASS	-
	Appearance	PASS	-
Test Device – Functionality	Mid positive	PASS	-
	High positive	PASS	-
	Negative 1	PASS	-
	Negative 2	PASS	-
Kit Components	Leak test - pouched device	PASS	-
	Leak test - diluent bottle	PASS	-

Kit Summary	Result
Total test devices assessed	15
Total test devices rejected	0
Kit status	PASS



GLOBAL DX LIMITED
 Eimbank Business Centre, Menstrie FK11 7BU
 United Kingdom
 Tel: +44 (0)1259 230830
 www.globaldx.com
 VAT No: GB 288 1975 44

Certified by:

Date:

21/12/2020

For in vitro diagnostic use only