



COAGDX31-1S v001

Certificate of Analysis

Product Information:

Product name:	Petscreen® Feline Parvovirus Ag (FPV) Test Kit
Product code:	GDX31-1S
Lot number:	GDX311S2001
Manufacture date:	June 2020
Expiry date:	May 2022

Quality Control Summary:

Component	Parameter	Result	Comments
Assay Diluent	Appearance	PASS	-
	Sterility	PASS	-
Test Device – Physical properties	Membrane flow	PASS	-
	Appearance	PASS	-
Test Device – Functionality	Low positive	PASS	-
	Mid positive	PASS	-
	High positive	PASS	-
	Negative 1	PASS	-
	Negative 2	PASS	-
	Negative 3	PASS	-
Kit Components	Leak test - pouched device	PASS	-
	Leak test - diluent bottle	PASS	-

Kit Summary	Result
Total test devices assessed	5
Total test devices rejected	0
Kit status	PASS



GLOBAL DX LIMITED
Elmbank Business Centre, Menzies FK11 7BU, United Kingdom
Tel: +44 (0)1259 230830
www.globaldx.com
VAT No: GB 256 1975 64

Certified by: [Signature] Date: 06/03/2020

For in vitro diagnostic use only